

**EU Declaration of Conformity**  
**according to Annex IV of Regulation (EU) 2017 /745 on medical devices**



We, the company

**Sameda GmbH**

Am Petersberg 36  
29389 Bad Bodenteich  
Germany  
SRN: DE-MF-000008334  
declare in sole responsibility,

that the brace listed below and described and specified in the technical documentation  
„TD 003-04 BETA-Flex Brace Fußabduktionsschiene“

**BETA-Flex Brace Fußabduktionsschienen - BETA-Flex Brace Foot Abduction Braces**  
**BETA-Flex Brace Fußabduktionsschiene - BETA-Flex Brace Foot Abduction Brace, BE 0600-000**

fulfils the general safety and performance requirements of

**Regulation (EU) 2017 /745 on medical devices (MDR), Annex 1.**

The BETA-Flex Brace Foot Abduction Brace is used for babies/toddlers and children diagnosed with idiopathic clubfoot according to the conservative therapy approach of Dr Ponseti. As one of the system components of the foot abduction orthosis it is used with foot supports.

**Basic-UDI-DI** 426020155FASZBSXV

**Risk class:** I  
(in accordance with MDR, Annex VIII) (Rule 1)

**Performed conformity assessment procedure:** MDR, Article 52 (7)

After additions or changes to the product or to the technical documentation the declaration of conformity is reissued.

Bad Bodenteich, 22<sup>nd</sup> March 2022  
Place, Date

  
  
Am Petersberg 36 • D-29389 Bad Bodenteich • Germany  
Beate Wicher 05824 98 555-0 • Fax.: +49 (0)5824 98 555 20  
general manager